

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK



UKAS  
CALIBRATION

0809

Accredited to  
ISO/IEC 17025:2005

### Seaward Electronic Ltd

Issue No: 009 Issue date: 06 July 2012

Bracken Hill  
South West Industrial Estate  
Peterlee  
County Durham  
SR8 2SW

Contact: Mr M Steele  
Tel: +44 (0) 191 586 3511  
Fax: +44 (0) 191 586 0227  
E-Mail: service@seaward.co.uk  
Website: www.seaward.co.uk

Calibration performed at the above address only

#### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
<b>ELECTRICAL CALIBRATION</b>			
<b>DC VOLTAGE</b>			
Generation	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	31 ppm + 3.7 μV 32 ppm + 4.8 μV 26 ppm + 45 μV 31 ppm + 0.46 mV 31 ppm + 4.6 mV	
Measurement	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	44 ppm + 2.9 μV 14 ppm + 1.2 μV 12 ppm + 4.3 μV 18 ppm + 46 μV 18 ppm + 0.23 mV	
<b>DC RESISTANCE</b>			
Generation	10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ 10 MΩ 100 MΩ 1 GΩ	6.9 mΩ 12 mΩ 94 mΩ 0.94 Ω 9.4 Ω 0.16 kΩ 4.6 kΩ 0.59 MΩ 14 MΩ	Nominal values obtained from a multi- function calibrator for calibration of multimeters, resistance meters etc.
	100 μΩ 1 mΩ 10 mΩ 100 mΩ 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ	20 ppm 15 ppm 8.0 ppm 7.6 ppm 6.9 ppm 6.6 ppm 13 ppm 7.7 ppm 10 ppm 6.5 ppm 7.9 ppm	



0809  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Seaward Electronic Ltd**  
**Issue No: 009 Issue date: 06 July 2012**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks
DC RESISTANCE (continued)			
Measurement	1 mΩ 10 mΩ 100 mΩ 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ	47 ppm 24 ppm 10 ppm 8.9 ppm 9.2 ppm 12 ppm 9.2 ppm 11 ppm 9.1 ppm 8.9 ppm	Calibration of standard resistors
	100 μΩ to 2 mΩ 2 mΩ to 20 mΩ 20 mΩ to 200 mΩ 200 mΩ to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 kΩ 2 kΩ to 20 kΩ 20 kΩ to 200 kΩ 200 kΩ to 2 MΩ 2 MΩ to 20 MΩ 20 MΩ to 200 MΩ 200 MΩ to 2 GΩ	46 ppm + 0.012 μΩ 22 ppm + 0.12 μΩ 9.9 ppm + 0.12 μΩ 8.9 ppm + 0.46 μΩ 9.2 ppm + 4.6 μΩ 12 ppm + 46 μΩ 9.2 ppm + 0.46 mΩ 11 ppm + 6.9 mΩ 9.0 ppm + 69 mΩ 8.9 ppm + 0.12 Ω 36 ppm + 860 Ω 0.035 % + 65 kΩ 0.23 % + 200 kΩ	
High resistance/high voltage Measurement	0.1 MΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 1 GΩ 1 GΩ to 10 GΩ 10 GΩ to 100 GΩ 100 GΩ to 1 TΩ	0.060 % 0.059 % 0.061 % 0.053 % 0.054 % 0.072 % 0.25 %	The applied voltages will be in the range 10 V to 1000 V DC
	1 MΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 1 GΩ 1 GΩ to 10 GΩ 10 GΩ to 100 GΩ 100 GΩ to 1 TΩ	0.32 % 0.31 % 0.31 % 0.31 % 0.37 % 0.44 %	



0809  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Seaward Electronic Ltd**  
**Issue No: 009 Issue date: 06 July 2012**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
AC VOLTAGE			
Measurement	20 mV to 200 mV 45 Hz to 55 Hz 55 Hz to 1 kHz	0.012 % + 45 $\mu$ V 0.012 % + 61 $\mu$ V	
	200 mV to 2 V 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 2 kHz 2 kHz to 10 kHz	92 ppm + 0.35 mV 70 ppm + 0.35 mV 74 ppm + 0.49 mV 92 ppm + 1.1 mV	
	2 V to 20 V 45 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz	93 ppm + 3.2 mV 77 ppm + 1.6 mV 93 ppm + 4.7 mV	
	20 V to 200 V 45 Hz to 100 Hz 100 Hz to 1 kHz	92 ppm + 33 mV 72 ppm + 33 mV	
	200 V to 1000 V 45 Hz to 100 Hz 100 Hz to 1 kHz	94 ppm + 0.27V 93 ppm + 0.78V	
Generation	20 mV to 200 mV 45 Hz to 1 kHz 1 kHz to 10 kHz	0.040 % + 43 $\mu$ V 0.083 % + 54 $\mu$ V	
	200 mV to 2 V 45 Hz to 1 kHz 1 kHz to 10 kHz	0.037 % + 0.11 mV 0.074 % + 0.16 mV	
	2 V to 20 V 45 Hz to 1 kHz 1 kHz to 10 kHz	0.035 % + 1.0 mV 0.064 % + 1.6 mV	
	20 V to 200 V 45 Hz to 1 kHz 1 kHz to 10 kHz	0.037 % + 8.7 mV 0.085 % + 46 mV	
	200 V to 1000 V 45 Hz to 1 kHz	0.037 % + 87 mV	Minimum frequency 46 Hz above 700 V



0809  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Seaward Electronic Ltd**  
**Issue No: 009 Issue date: 06 July 2012**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
<b>DC CURRENT</b>			
Generation	0 $\mu$ A to 200 $\mu$ A 200 $\mu$ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 10 A 10 A to 20 A	94 ppm + 37 nA 83 ppm + 68 nA 57 ppm + 0.69 $\mu$ A 120 ppm + 3.6 $\mu$ A 0.015 % + 40 $\mu$ A 0.035 % + 0.49 mA 0.035 % + 0.79 mA	
Calibration of clamp-on ammeters using multi-turn coil	20 A to 100 A 100 A to 1000 A	0.55 % + 0.48 A 0.33 % + 46 mA	Hall effect clamps Wound clamps
Measurement	0 $\mu$ A to 200 $\mu$ A 200 $\mu$ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	29 ppm + 25 nA 29 ppm + 0.11 $\mu$ A 29 ppm + 1.0 $\mu$ A 58 ppm + 14 $\mu$ A 170 ppm + 0.33 mA	
<b>AC CURRENT</b>			
Generation	<i>45 Hz to 500 Hz:</i> 20 $\mu$ A to 200 $\mu$ A 200 $\mu$ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A  <i>45 Hz to 100 Hz:</i> 2 A to 20 A	0.12 % + 0.50 $\mu$ A 0.087 % + 0.61 $\mu$ A 0.086 % + 4.6 $\mu$ A 0.086 % + 46 $\mu$ A 0.098 % + 0.46 mA  0.060 % + 2.9 mA	
Calibration of clamp-on ammeters using multi-turn coil	<i>40 Hz to 60 Hz:</i> 20 A to 1000 A 20 A to 1000 A	0.55 % + 0.48 A 0.33 % + 46 mA	Hall effect clamps Wound clamps
Measurement	<i>50 Hz to 200 Hz:</i> 2 $\mu$ A to 200 $\mu$ A 200 $\mu$ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	0.024 % + 330 nA 0.026 % + 1.8 $\mu$ A 0.024 % + 35 $\mu$ A 0.024 % + 180 $\mu$ A 0.059 % + 1.7 mA	
<b>FREQUENCY</b>			
	1 Hz to 10 Hz 10 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 10 MHz 10 MHz to 100 MHz 100 MHz to 1 GHz	11 ppm + (0.62 mHz to 1.3 mHz) 11 ppm + 1.3 mHz 11 ppm + (1.3 mHz to 1.2 mHz) 11 ppm + 1.2 mHz 11 ppm + (1.2 mHz to 1.3 mHz) 11 ppm + (1.3 mHz to 5.9 mHz) 11 ppm + (5.9 mHz to 58 mHz) 11 ppm + (58 mHz to 0.58 Hz) 11 ppm + (0.58 Hz to 5.8 Hz)	



0809  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Seaward Electronic Ltd**  
**Issue No: 009 Issue date: 06 July 2012**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
<b>Calibration of Portable Appliance Testers</b>			
Earth Bond	<i>At 50 Hz:</i> 0.05 $\Omega$ 0.09 $\Omega$ 0.1 $\Omega$ 0.115 $\Omega$ 0.2 $\Omega$ 0.33 $\Omega$ 0.45 $\Omega$ 0.5 $\Omega$ 0.575 $\Omega$ 1.0 $\Omega$ 3.3 $\Omega$ 5.0 $\Omega$ 10 $\Omega$ 15 $\Omega$ 18 $\Omega$	1.2 % 0.67 % 0.62 % 0.54 % 0.34 % 0.30 % 0.28 % 0.24 % 0.25 % 0.20 % 0.29 % 0.24 % 0.19 % 0.18 % 0.18 %	
Insulation Resistance	<i>At DC:</i> 0.25 M $\Omega$ 0.5 M $\Omega$ 1.0 M $\Omega$ 2.0 M $\Omega$ 4.0 M $\Omega$ 5.0 M $\Omega$ 6.0 M $\Omega$ 10 M $\Omega$ 15 M $\Omega$ 18 M $\Omega$ 50 M $\Omega$ 90 M $\Omega$	0.25 % 0.13 % 0.077 % 0.060 % 0.062 % 0.066 % 0.052 % 0.051 % 0.068 % 0.064 % 0.083 % 0.12 %	
Insulation Resistance Test Voltage	100 V 250 V 500 V 1000 V	0.094 % 0.073 % 0.13 % 0.11 %	
Leakage Current	<i>DC:</i> 0 $\mu$ A to 320 $\mu$ A 320 $\mu$ A to 3.2 mA 3 mA to 32 mA  <i>50 Hz to 60 Hz:</i> 0 mA to 1.6 mA 1.6 mA to 16 mA 16 mA to 20 mA	0.12 % + 0.051 $\mu$ A 0.069 % + 0.23 $\mu$ A 0.11 % + 2.3 $\mu$ A  0.29 % + 0.40 $\mu$ A 0.35 % + 3.8 $\mu$ A 0.35 % + 39 $\mu$ A	



0809  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Seaward Electronic Ltd**  
**Issue No: 009 Issue date: 06 July 2012**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
Calibration of Portable Appliance Testers (continued)			
Flash Test	<i>DC, 50 Hz and 60 Hz:</i>  Current (0 mA to 32 mA)  Voltage (AC, 50 Hz and 60 Hz) 1 kV 1.5 kV 2 kV 3 kV 4 kV 5 kV  Voltage (DC) 1 kV 2 kV 3 kV 4 kV 5 kV 6 kV	See <i>Leakage Current</i> above  12 V 13 V 17 V 21 V 28 V 35 V  10 V 12 V 13 V 14 V 20 V 32 V	These capabilities also apply to flash testers and high voltage meters.
Load Test	<i>At 50 Hz:</i> 1 kVA and 3 kVA nominal	1.5 % + 0.015 kVA	
<b>Calibration of RCD Testers</b>			
Trip Current	<i>At 50 Hz:</i> 10 mA to 100 mA 100 mA to 2 A	1.4 % + 0.083 mA 1.4 % + 0.83 mA	
Trip Time	20 ms to 100 ms 100 ms to 400 ms 400 ms to 700 ms 700 ms to 900 ms	0.96 ms 1.5 ms 4.8 ms 8.5 ms	
<b>Calibration of Loop Testers</b>			
Loop impedance	<i>At 50 Hz:</i> 0.5 $\Omega$ to 1 k $\Omega$	(0.60 % to 0.80 %) + 4.6 m $\Omega$	



0809  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Seaward Electronic Ltd**  
**Issue No: 009 Issue date: 06 July 2012**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks		
<b>CALIBRATION OF SEAWARD PV100/PV150 SOLAR INSTALLATION METERS</b>					
Insulation Resistance	0.5 M $\Omega$ 1 M $\Omega$ 10 M $\Omega$ 90 M $\Omega$ 190 M $\Omega$	10 k $\Omega$ 10 k $\Omega$ 100 k $\Omega$ 100 k $\Omega$ 1.0 M $\Omega$	At 500 V DC		
	0.25 M $\Omega$ 90 M $\Omega$	10 k $\Omega$ 1.0 M $\Omega$			
	1 M $\Omega$ 90 M $\Omega$	10 k $\Omega$ 1.0 M $\Omega$			
Insulation Voltage	250 V 500 V	1.0 V 1.0 V	Into 1 mA load Into 1 mA load		
	Continuity Resistance	0.67 $\Omega$ 1.2 $\Omega$ 10.5 $\Omega$ 100 $\Omega$ 190 $\Omega$		10 m $\Omega$ 10 m $\Omega$ 100 m $\Omega$ 1.0 $\Omega$ 1.0 $\Omega$	
Current clamp input	100 mV (10 A simulation) 199 mV (20 A simulation) 500 mV (50 A simulation)	100 mA 100 mA 100 mA	Simulation of current clamp at 100 A/V		
RPE Voltage	110 V 240 V 420 V	1.0 V 1.0 V 1.0 V			
Open circuit panel voltage	10.9 V 49.9 V 202 V 440 V 640 V 900 V	100 mV 100 mV 1.0 V 1.0 V 1.0 V 1.0 V	DC voltage		
	Short circuit current	0.97 A 2.7 A 8 A		10 mA 20 mA 50 mA	DC Current



0809  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Seaward Electronic Ltd**  
Issue No: 009 Issue date: 06 July 2012

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	
<b>CALIBRATION OF DEFIBRILLATOR TESTERS</b>				
Energy	20 J to 70 J 70 J to 100 J 100 J to 360 J	1.5 % to 1.3 % 1.3 % to 1.2 % 1.2 %	By electrical simulation	
<b>CALIBRATION OF TEMPERATURE INDICATORS</b>				
PT100 indicators	- 125 °C to 0 °C 0 °C to 130 °C 130 °C to 266 °C 266 °C to 408 °C 408 °C to 558 °C 558 °C to 715 °C 715 °C to 883 °C	0.0025 °C to 0.0047 °C 0.0047 °C to 0.0070 °C 0.0070 °C to 0.0096 °C 0.0096 °C to 0.012 °C 0.012 °C to 0.015 °C 0.015 °C to 0.018 °C 0.018 °C to 0.021 °C		
PT25 indicators	5 °C to 408 °C 408 °C to 883 °C	0.0056 °C to 0.013 °C 0.013 °C to 0.024 °C		
Type K thermocouple indicators	- 190 °C to - 154 °C - 154 °C to - 115 °C - 115 °C to - 82 °C - 82 °C to - 53 °C - 53 °C to - 26 °C - 26 °C to 0 °C 0 °C to 25 °C 25 °C to 122 °C 122 °C to 246 °C 246 °C to 1370 °C	0.26 °C to 0.20 °C 0.20 °C to 0.17 °C 0.17 °C to 0.16 °C 0.16 °C 0.15 °C 0.15 °C 0.15 °C 0.15 °C to 0.14 °C 0.14 °C to 0.16 °C 0.16 °C		
<b>PRESSURE CALIBRATION</b>				
Gas Pressure, Gauge	- 85 kPa to 0.0 kPa 0.0 kPa to 50 kPa 50 kPa to 1.0 MPa 1.0 MPa to 2.0 MPa	1.0 kPa 24 Pa 1.1 kPa 1.0 kPa		
END				





0809  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Seaward Electronic Ltd**  
Issue No: 009 Issue date: 06 July 2012

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

*A CMC is a calibration and measurement capability available to customers under normal conditions:*

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or*
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.*

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

- As a single value that is valid throughout the range.
- As an explicit function of the measurand or of a parameter (see below).
- As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.
- As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.
- In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

**Expression of CMCs - symbols and units**

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples, and an indication of how they are to be interpreted, are shown below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0  $\mu$ V:

Over the range 100 mV to 1 V, the CMC is 0.0025 %  $\cdot$  V + 5.0  $\mu$ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %  $\cdot$  p + (0.12  $\cdot$  10<sup>-6</sup>  $\cdot$  p  $\cdot$  10<sup>-6</sup>) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5  $\cdot$  0.01  $\cdot$  i, where i is the instrument indication.